

RESEARCH ARTICLE

Open Access

Impact of delays to incubation and storage temperature on blood culture results: a multi-centre study



Clare L. Ling^{1,2*} , Tamalee Roberts³, Sona Soeng⁴, Tomas-Paul Cusack^{3,5}, David A. B. Dance^{2,3,6}, Sue J. Lee^{2,7}, Thomas A. N. Reed^{4,8}, Pattaraporn Hinfonthong¹, Somsavanh Sihalath³, Amphone Sengduangphachanh³, Wanitda Watthanaworawit¹, Tri Wangrangsimakul^{2,7}, Paul N. Newton^{2,3,6,7}, Francois H. Nosten^{1,2}, Paul Turner^{2,4} and Elizabeth A. Ashley^{2,3,9}

Abstract

Background: Blood cultures are one of the most important tests performed by microbiology laboratories. Many hospitals, particularly in low and middle-income countries, lack either microbiology services or staff to provide 24 h services resulting in delays to blood culture incubation. There is insufficient guidance on how to transport/store blood cultures if delays before incubation are unavoidable, particularly if ambient temperatures are high. This study set out to address this knowledge gap.

Methods: In three South East Asian countries, four different blood culture systems (two manual and two automated) were used to test blood cultures spiked with five common bacterial pathogens. Prior to incubation the spiked blood culture bottles were stored at different temperatures (25 °C, in a cool-box at ambient temperature, or at 40 °C) for different lengths of time (0 h, 6 h, 12 h or 24 h). The impacts of these different storage conditions on positive blood culture yield and on time to positivity were examined.

Results: There was no significant loss in yield when blood cultures were stored < 24 h at 25 °C, however, storage for 24 h at 40 °C decreased yields and longer storage times increased times to detection.

Conclusion: Blood cultures should be incubated with minimal delay to maximize pathogen recovery and timely result reporting, however, this study provides some reassurance that unavoidable delays can be managed to minimize negative impacts. If delays to incubation \geq 12 h are unavoidable, transportation at a temperature not exceeding 25 °C, and blind sub-cultures prior to incubation should be considered.

Keywords: Blood cultures, Delays, LMIC, Diagnostics, Microbiology

* Correspondence: clare@tropmedres.ac

¹Shoklo Malaria Research Unit, Mahidol-Oxford Tropical Medicine Research Unit, Faculty of Tropical Medicine, Mahidol University, Mae Sot, Thailand

²Centre for Tropical Medicine and Global Health, Nuffield Department of Medicine, University of Oxford, Oxford, United Kingdom

Full list of author information is available at the end of the article



© The Author(s). 2021 **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

Background

Blood cultures (BCs) are one of the most important tests performed by microbiology laboratories. Usually taken from the most unwell patients, they may reveal the likely source of infection and guide antibiotic therapy. Moreover, they can provide reliable surveillance data for informing national antibiotic treatment guidelines. The Global Antimicrobial Resistance Surveillance System (GLASS) of the World Health Organization identifies BCs as key for antimicrobial resistance surveillance in all Member States and many low and middle-income countries (LMICs) are increasing their laboratory capacity to meet this target [1].

Several initiatives have been evaluated to reduce the time to identify and test antimicrobial susceptibility of pathogens causing bacteremia with the joint goals of improving patient management and antimicrobial stewardship, for example around-the-clock processing and direct antimicrobial susceptibility testing from positive BC bottles [2, 3]. However, all these attempts can be undermined by transportation delays or other factors leading to delays in incubating BCs. Such delays occur when laboratories do not operate a 24 h service, which is common in LMICs or when BCs are referred from other hospitals to a centralized laboratory, also common in LMICs and increasingly in other settings due to consolidation of pathology services [4, 5]. Delayed incubation leads not only to delayed pathogen identification but may also affect the diagnostic yield of BCs [6, 7]. In addition, in some LMICs BCs are likely to be stored at tropical ambient temperatures that can exceed recommended incubation temperatures.

This study investigated the effects of storage for different times at different temperatures on the isolation of some important bacteria frequently grown in BCs. To increase the generalizability of the results, the study was performed in three South East Asian countries using two different manual and two different automated BC systems. Pediatric bacteremia was simulated, with effects on yield and time to positivity (TTP) explored.

Methods

Sites

The study took place in laboratories at: the Cambodia Oxford Medical Research Unit (COMRU), Siem Reap, Cambodia; the Shoklo Malaria Research Unit (SMRU), Tak Province, Thailand; and the Lao-Oxford-Mahosot Hospital-Wellcome Trust Research Unit (LOMWRU), Vientiane, Lao PDR.

Blood cultures

Manual BC systems were used at COMRU and LOMWRU, and automated BC systems were used at LOMWRU and SMRU. All sites used pediatric aerobic BC

bottles. At COMRU, BC bottles contained 25 ml of Brain Heart Infusion broth plus 0.025% sodium polyanethol sulfonate (SPS); at LOMWRU, BC bottles for their manual method contained 20 ml tryptic hydrolysate of casein and soy peptone broth (TSB) with 0.05% SPS while PEDS Plus™/F bottles (Becton Dickinson, New Jersey, USA) were used with the automated BD BACTEC™ system and at SMRU, PF Plus bottles were used with the automated BacT/ALERT® system.

BC bottles were inoculated with 2 ml fresh untreated human blood intended to contain approximately 3 CFU/ml of one of five bacterial strains: *Escherichia coli* (ATCC 25922), *Staphylococcus aureus* (ATCC 25923), *Streptococcus pneumoniae* (ATCC 49619), *Streptococcus agalactiae* (ATCC 12386), or *Haemophilus influenzae* (ATCC 10211). This was achieved by preparing suspensions of each control strain (using less than 24 h old bacterial cultures on solid media) in normal saline equivalent to a 1 McFarland standard (approximately 3×10^8 CFU/ml), further diluting them to approximately 30 CFU/ml with normal saline, and then making a final 1 in 10 dilution with donor blood to obtain a final concentration of approximately 3 CFU/ml. The Miles and Misra method was used to check actual concentrations [8]. The untreated blood was obtained from 13 healthy human volunteers (four each at SMRU and LOMWRU and five at COMRU). All volunteers were > 18 years of age, were not pregnant, had not previously been treated for anemia, and had not knowingly received antibiotics in the previous month. To avoid autoagglutination, volunteer blood was used immediately after collection, with time to bottle inoculation taking < 30 min.

Inoculated bottles were stored at different temperatures and times, with five replicates for each organism/storage condition combination, except LOMWRU where BACTEC bottles were tested in triplicate due to machine capacity (Table 1). Cool-boxes (double-skinned metal boxes) without icepacks, were kept in the shade at ambient temperatures at each site with their internal temperature monitored continuously with TinyTag Transit 2 (Gemini Data Loggers, UK) or MicroLite (Fourtec – Fourier Technologies Ltd., Australia) data loggers.

For BacT/ALERT® PF Plus bottles the color of the bottle sensor, which changes from green to yellow to indicate growth, was recorded pre-incubation. The bottles in the automated systems were incubated until they flagged positive or for at least 5 days. Sub-cultures were performed on bottles that flagged positive to confirm the identification of any viable bacteria and to rule out contamination. For manual BC systems, the bottles were inspected for visible growth and sub-cultured every 24 h until a positive sub-culture was obtained or for a maximum of 7 days.

Table 1 Storage conditions and number replicates for blood culture bottles spiked with five different organisms

Storage condition	Per organism					Total
	COMRU ^a	LOMWRU ^a	SMRU ^b	LOMWRU ^b	Total	
No storage	×5	×5	×5	×3	×18	90
6 h 25 °C	×5	×5	×5	×3	×18	90
6 h cool-box	×5	×5	×5	×3	×18	90
6 h 40 °C	×5	×5	×5	×3	×18	90
12 h 25 °C	×5	×5	×5	×3	×18	90
12 h cool-box	×5	×5	×5	×3	×18	90
12 h 40 °C	×5	×5	×5	×3	×18	90
24 h 25 °C	×5	×5	×5	×3	×18	90
24 h cool-box	×5	×5	×5	×3	×18	90
24 h 40 °C	×5	×5	×5	×3	×18	90
Total	50	50	50	30	180	900

^a = manual system; ^b = automated system (BacT/ALERT[®] for SMRU, BD BACTEC[™] for LOMWRU)

Statistical analysis

The percentage of positive cultures detected was summarized by storage condition. Pairwise differences across groups were compared using Fisher's exact test. The sample size of 90 bottles for each storage condition enabled the detection of a difference in the proportion of culture growth from 89 to 73%, assuming an alpha of 0.05 and 80% power. As a secondary outcome, median TTP (p25, p75) from inoculation was summarized for each storage condition and compared using the Wilcoxon rank-sum test and a test-for-trend across groups. *P* values < 0.05 were considered statistically significant. Analyses were performed using STATA version 15.1 (StataCorp LLC, USA).

Results

Inoculum

Using the Miles and Misra method the bottles were confirmed to have been inoculated with blood containing 1 to 10 CFU/ml (median 2.9 CFU/ml) of bacteria, resulting in 2 to 20 CFU/bottle. For *E. coli*, *H. influenzae*, *S. aureus*, *S. agalactiae* and *S. pneumoniae* the median CFU/ml were 2.3 (range 2.0–3.0), 4.5 (range 3.2–7.0), 4.1 (range 2.3–7.0), 2.2 (range 1.7–2.9) and 1.7 (range 1.0–10.0), respectively.

Sub-culture results

For each organism (*E. coli*, *H. influenzae*, *S. agalactiae*, *S. aureus* and *S. pneumoniae*) 180 bottles were inoculated of which 80.6, 95.0, 90.0, 93.3 and 66.1% were positive respectively. The expected organism was obtained from all positive bottles when they were sub-cultured except for one bottle inoculated with *S. pneumoniae* but grew a *Staphylococcus* species; this bottle was regarded as contaminated and excluded from further analysis. In addition, at SMRU, 19 bottles inoculated with *S.*

pneumoniae flagged positive in the Bact/Alert system, but failed to grow an organism when sub-cultured. These 19 'positive' bottles showed exponential growth on their charts, BC bottle media showed evidence of browning, Gram positive cocci were seen on Gram staining and their median time from flagging positive to sub-culture (13.73 h, IQR 10.05–14.65) was significantly longer than the median time taken to sub-culture the 23 bottles at SMRU inoculated with *S. pneumoniae* that did grow (7.57 h, IQR 0.43–12.07, *P* < 0.001). It was concluded that failure of sub-culture for these 19 bottles was due to autolysis and to avoid bias due to sub-culture delay for this organism, they were included in the analysis as positive bottles.

Positive blood culture yield

The numbers of positive BC bottles obtained for different storage conditions are shown in Table 2.

When combining results for all organisms, no significant differences in positive bottle yields compared with no delay were detected when bottles were stored for 6 h, 12 h, at 25 °C or in a cool-box. The median (IQR; range) temperature inside the cool-boxes during the study was 27.0 °C (26.0–27.9, 19.8–31.0).

Significantly lower yields compared with no delay were detected when bottles were stored for 24 h or at 40 °C; and the lowest yield was obtained for bottles stored for 24 h at 40 °C. Storage for 24 h at 40 °C gave the lowest positive yield for both manual and automated systems [35/50 (70.0%) and 24/40 (60.0%), respectively].

When stratified by organism, no significant differences for any storage condition compared with no delay were detected for *E. coli*, *H. influenzae*, and *S. aureus*. For *S. agalactiae*, fewer bottles were positive when stored for 24 h at 40 °C (66.7%) and 24 h in a cool-box (55.5%) compared to no storage at all (100%, *P* = 0.019 and 0.003

Table 2 Overall number of positive blood cultures obtained for different storage conditions, compared with no delay; and number of BacT/ALERT® bottle sensors that indicated growth prior to incubation

Time (h)	Temp (°C)	Bottles (n)	Positive (n) ^a	Positive (%)	P value ^c	Sensor positive ^d
0	NA	90	80	88.9	NA	0/25
6	25 °C	90	85	94.4	0.281	0/25
6	Cool-box	90	81	90.0	1.000	0/25
6	40 °C	90	74	82.2	0.289	0/25
12	25 °C	90	76	84.4	0.511	0/25
12	Cool-box	89 ^b	80	89.9	1.000	0/25
12	40 °C	90	77	85.6	0.656	15/25
24	25 °C	90	82	91.1	0.805	20/25
24	Cool-box	90	71	78.9	0.104	20/25
24	40 °C	90	59	65.6	< 0.001	20/25
6	All	270	240	88.9	1.000	0/75
12	All	269 ^b	233	86.6	0.787	15/75
24	All	270	212	78.5	0.030	60/75
All	25 °C	270	243	90.0	0.841	20/75
All	Cool-box	269 ^b	232	86.2	0.694	20/75
All	40 °C	270	210	77.8	0.021	35/75

^aFor automated systems this is the number of bottles that flagged positive, and for manual systems this is the number of bottles that were positive on sub-culture; ^bOne bottle excluded due to contamination; ^cCompared with no delay; ^dSensor positive if BacT/ALERT® bottle sensor indicated growth prior to incubation. NA Not applicable

respectively). For *S. pneumoniae*, storage for 24 h at 40 °C gave a significantly lower number of positive bottles compared with no delay (27.8 and 72.2% respectively, $P = 0.018$). When comparing the positive yields obtained between the different systems, a significant difference was detected between the two manual methods ($P = 0.037$).

Time to positivity

The overall median TTP from inoculation for BCs performed with the automated systems was 18.9 h (IQR 15.6–26.0), with *E. coli* having the shortest TTP (15.9, IQR 14.2–25.2). Both *H. influenzae* (22.8, IQR 17.7–28.0) and *S. aureus* (22.1, IQR 17.6–28.0) had significantly longer TTP from inoculation than the other three bacterial species ($P < 0.001$ – 0.018). The overall median TTP increased from 13.0 h (IQR 12.3–16.4) with no delay, to 28.1 h (IQR 26.8–30.9) after 24 h of storage (P for trend < 0.001) (Fig. 1). There was no significant difference in the TTP between the two automated BC systems.

For manual BC systems the majority of positive sub-cultures was obtained for the first 24 h sub-culture (393/409, 96.1%) and no new positive sub-culture was obtained after 4 days of incubation. The 16 positive bottles that had positive sub-cultures following > 24 h incubation had been stored at 25 °C for 24 h (eight bottles); in a cool-box for 24 h (five bottles); in a cool-box for 12 h

(two bottles) or at 25 °C for 6 h (1 bottle). Of the 16 bottles that had positive sub-cultures following > 24 h incubation, two had been cultured at LOMWRU and 14 had been cultured at COMRU (0.9 and 7.2% of positive bottles at these sites respectively, $P = 0.001$).

Visual detection of growth

Detection of growth by visual inspection was recorded daily for manual methods (with the exception of *S. pneumoniae* at COMRU, which was omitted in error). There was discordance in both directions when comparing visual detection of growth with detection of growth obtained by sub-culturing, however, overall visual detection significantly underestimated the number of actual sub-culture positive bottles with 288/450 (64%) versus 362/450 (80.4%) determined to be positive respectively ($P < 0.001$).

Visual detection of growth prior to incubation is possible for BacT/ALERT® PF Plus bottles that have a colorimetric growth sensor. Recording of BacT/ALERT® bottle sensor colours showed growth before incubation increased with increased time to incubation (Table 2). Fewer bottles that indicated growth prior to incubation flagged positive (62/75, 82.7%) compared to those that did not indicate growth prior to incubation (164/175, 93.7%) ($P = 0.01$). Of the 13 bottles with positive growth indicators that failed to flag positive, 12 had been stored at 40 °C for ≥ 12 h.




Fig. 1 Time to positivity from time of inoculation for automated blood cultures. The box graph shows the time to positivity (TTP) for automated blood cultures for different storage conditions

Discussion

There are many factors that can affect pathogen recovery from BCs, from the processes of blood collection through to the methodologies used in the laboratory. There are guidelines describing best practice, but these do not sufficiently address how to manage delayed incubation, particularly in tropical climates. A review describing the best practice for BCs in LMICs highlights the issue of delayed incubation, however, optimal conditions have not been established [9]. The UK Standard for Microbiology Investigations for BCs recommends that bottles should be placed in incubators no later than 4 h after inoculation [10]. This is not always possible; for example the laboratory in Laos receives BCs from > 300 km away, and only once per day. This study set out to determine the impact of such, sometimes unavoidable, delays on BC results.

S. pneumoniae autolysis

In this study when using the BacT/Alert system, pneumococci were recovered from 11/11 of positive BC bottles sub-cultured < 7 h from the time they flagged positive and just 12/31 when sub-cultured after 7 h. The reduction in recovery was most likely due to autolysis,

which has been described previously for the BacT/Alert automated system with similar recovery rates, and the continuous agitation speculated to contribute to the phenomenon [11–15]. This finding supports recommendations that positive BC bottles in automated systems should be sub-cultured immediately to minimize the risk of false negative results and reporting delays [16], with implications for laboratories that are not staffed 24 h/7 days a week, a common occurrence in LMIC settings.

Impact of delayed incubation on yield

The main aim of this study was to assess the impact of delayed incubation on positive BC bottle yield. It is encouraging that in this study no significant differences in positivity compared with no delay were obtained when inoculated bottles were stored for 6 h or 12 h irrespective of the storage temperature. When storing bottles for 24 h the temperature was found to be important with the lowest yield of 65.6% obtained for bottles stored at 40 °C; compared with 78.9 and 91.1% when stored in a cool-box or at 25 °C respectively. Ambient temperature is usually recommended for transportation of BCs; however, the results presented here indicate that this may be suboptimal in tropical climates and bottles should be

protected from high temperatures where possible. This agrees with a review on how to optimize BCs that stated that pre-incubation at 35 °C should be discouraged [17]. In addition, a previous study comparing delayed incubation at room temperature with 35 °C, found positive BC yields decreased with increasing delay time more in BCs stored at 35 °C compared with room temperature [18].

To provide a relatively protected environment at ambient temperatures, the use of cool-boxes without ice packs offers a cost-effective option, and promisingly the positivity rates in this study were higher for samples held in cool-boxes as opposed to 40 °C. Nevertheless, some caution must be used when taking this approach, for example during periods of very high ambient temperature or when exposure of the boxes to direct sunlight cannot be avoided. The highest recorded temperature inside the cool-boxes was 31.0 °C, but temperatures > 40 °C have been recorded previously [19] and were observed in this study when temperature monitoring continued (Table S2, Figures S1, S2 and S3). Further studies into the utility of cool-boxes in tropical climates with and without ice packs are warranted as this is a practicable method of storage and transportation for samples in LMICs.

Impact of delayed incubation on time to positivity

Although delays before incubation did not affect the proportion of positive BCs if the temperature was controlled, they did affect the TTP from bottle inoculation. This was illustrated best by the automated systems where the median TTP increased from 13.0 h (IQR 12.3–16.4) with no delay to 28.1 h (IQR 26.8–30.9) following 24 h of storage (P for trend < 0.001) (Fig. 1). This finding agrees with a previous study that demonstrated TTP increased for samples kept at room temperature for 19 h compared with 2 h or 9.5 h (P < 0.001) [20]. This has implications for result reporting and patient management, with potential negative impacts on patient outcome if inadequate empiric treatment has been used. For example, a previous study found TTP was inversely associated with mortality risk for patients with Gram negative bacteremia [21]. This provides further evidence that delays to incubation should be avoided where possible.

Visual detection of growth

When processing manual BCs, reliance on visual detection of growth is dependent on local standard operating procedures. The accuracy of visual detection of growth is multifactorial and may vary between individuals [9]. In this study visual detection of growth significantly underestimated the the number of positive bottles detected by sub-culture (P < 0.001), illustrating that visual detection of growth should not be relied upon.

Recording of BacT/ALERT® bottle sensor colours prior to incubation showed growth before incubation increased

with increased time to incubation and that fewer bottles that indicated growth prior to incubation flagged positive compared to those that did not indicate growth prior to incubation (P = 0.01). Of the 13 bottles with positive growth indicators that failed to flag positive, 12 had been stored at 40 °C for \geq 12 h. Failure of bottles to flag positive despite indicating growth prior to incubation may have been due to loss of viability and/or due to insufficient changes in the colorimetric sensor for the machine to detect (exhaustive CO₂ production prior to incubation).

Utility of blind sub-cultures

Results from this study including delayed TTP, failure of positive bottles to flag positive in the automated BC system and unreliability of visual checks for growth for the manual systems indicate that a blind sub-culture before incubation may help to reduce false negative results and decrease TTP from inoculation if BC bottles are unavoidably stored for extended periods of time. Previous studies evaluating the utility of blind sub-cultures for diagnostic BCs concluded that early blind sub-cultures were useful, with one study recommending performing a blind sub-culture 12–17 h following inoculation and one 4–14 h following arrival in the laboratory [22–24]. In practice, the utility, practicability and timing of an early blind sub-culture will depend on laboratory and biological factors including working hours, staffing levels, workload, blood volumes and bacterial loads. The majority of manual culture bottles were positive for their first sub-culture performed at 24 h (96%), further supporting the use of blind sub-cultures, with at least one being performed following 24 h of incubation. This is in agreement with previous studies and guidelines [10, 25].

Automated versus manual blood cultures

In this study, both manual and automated systems were included; these reflected the systems in-use at the different sites at the time. Overall, the automated systems performed better than the manual systems resulting in significantly more positive BC bottles (356/399, 89.2% versus 409/500, 81% respectively; P = 0.002). In addition to other benefits, such as continuous-monitoring to enable positive cultures to be detected sooner and savings of staff time, automated systems have previously been reported to have increased sensitivity and decreased time to positivity compared with manual BCs [26]. This finding supports the use and implementation of automated BC bottles in all settings, where financially viable. Significant differences were detected when yield and TTP from inoculation were compared for the two manual BC systems that may have resulted from the different broths used, highlighting the advantage of standardized automated BC bottles.

Study design and limitations

As previously shown, low-colony-count neonatal sepsis can occur [27, 28]. In this study a 2 ml inoculum containing approximately 3 CFU/ml was chosen to represent the lowest concentration suitable to conduct the experiments. Although the study was designed to ensure the results were generalizable, it did have some limitations. When stratified by organism, storage conditions and BC system, the sample sizes were small. Different healthy volunteers donated blood, opposed to infected individuals with unknown impact. The experiments were challenging to perform and for practical reasons a limited pathogen/strain range was used, with only one blood inoculum volume. In this study, the recovery of *Streptococcus* spp. were more susceptible to the adverse effects of storage. In reality, the pathogen range seen in routine practice is much greater, and organisms encountered, particularly as part of a natural infection, may be affected by different storage conditions to a lesser or greater extent than described here. In addition, the impact of delays to incubation on polymicrobial infections were not investigated. It is possible that the presence of contaminating organisms could overgrow true pathogens with longer delays to incubation.

Variation in results was observed between the different sites and for the different organisms. This was anticipated, however, the results for some organisms varied more than expected. For example, the positive yield of *E. coli* was just 36% (18/50) at COMRU compared with 97.7% (127/130) at other sites (Table S1). Differing inocula would not explain the variation seen for *E. coli* (range 2–3 CFU/ml), and the same batch of media used for the *E. coli* experiment supported the growth of more fastidious organisms. It is therefore probable that there was a technical issue for this experiment, for example inadequate mixing or an interaction between the bacteria and the donor blood.

Conclusions

The results from this study, together with findings from other studies, lead to the following recommendations:

- BC storage time should be minimized where possible (preferably no longer than 12 h);
- if storage for longer than 12 h is unavoidable, BC bottles should be protected from high temperatures, for example by storing them in a cool-box (maintaining a temperature of approximately 25 °C was shown to be suitable in this study) away from direct sunlight;
- a blind sub-culture prior to incubation on all BC bottles that have had prolonged storage should be

performed (preferably on all bottles that have been stored for ≥ 12 h);

- for manual BC systems, visual inspection alone should not be relied upon and at least one sub-culture following 24 h incubation should be performed; and
- for automated BC systems, delays in sub-culturing bottles that have flagged positive should be minimized (this and a previous study support a maximum delay of 7 h to maximize the recovery of *S. pneumoniae*).

This study provides useful insight into the impact of different storage conditions prior to incubation on BC results. Although improved access to microbiology services in LMICs is the aspiration, this study provides some reassurance that unavoidable delays in incubating BCs can be managed to minimize any negative impacts.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s12879-021-05872-8>.

Additional file 1: Table S1. Experiment details and numbers of positive blood culture bottles, irrelevant of storage condition, per organism.

Tables S2-S6. Number of positive blood culture bottles for different storage conditions for each organism. **Tables S7.** Cool box temperatures during and following the study period. **Figure S1-S3.** Line graphs of cool box temperatures obtained during and following the study period for LOMWRU, COMRU and SMRU.

Abbreviations

COMRU: Cambodia Oxford Research Unit; LMICs: Low and middle-income countries; LOMWRU: Lao-Oxford-Mahosot Hospital-Wellcome Trust Research Unit; SMRU: Shoklo Malaria Research Unit; TTP: Time to positivity

Acknowledgements

We thank all the blood donor volunteers and laboratory staff involved in this study. In addition we thank Cecilia Ferreyra, Elisa Baszanger and Cassandra Kelly from FIND for their support of this project.

Authors' contributions

CLL, TR, DABD, SJL, TPC, TW, PT and EAA designed the study. TR, SS, TANR, TPC, PH, SS and AS performed the experiments with supervision from CLL, DABD, WW, PNN, FHN, PT and EAA. CLL and SJL analysed the data. All authors were involved in manuscript writing and review; and all authors read and approved the final manuscript.

Funding

This work was supported by the Foundation for Innovative New Diagnostics (FIND), Switzerland and by the Wellcome Trust, UK. FIND reviewed and approved the study design.

Availability of data and materials

The dataset supporting the conclusions of this article is available in the Oxford University Research Archive (ORA) repository (<https://ora.ox.ac.uk/objects/uuid:8c023bb7-496c-4926-a556-f38d0c289c4e>) (DOI: <https://doi.org/10.5287/bodleian:NGO7xqboq>).

Ethics approval and consent to participate

All three study sites (COMRU, SMRU and LOMWRU) are part of the Oxford Tropical Network and the study was approved by the Oxford Tropical Ethics Committee (OxTREC reference number 35–17). In addition, the Angkor Hospital for Children's Research Committee approved the study (Reference

number 1236/18AHC). Local ethical approval was not obtained in Thailand and Laos, as the study was deemed to be of low risk and human volunteers were restricted to staff members. All human volunteers in this study provided written informed consent.

Consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Author details

¹Shoklo Malaria Research Unit, Mahidol-Oxford Tropical Medicine Research Unit, Faculty of Tropical Medicine, Mahidol University, Mae Sot, Thailand.

²Centre for Tropical Medicine and Global Health, Nuffield Department of Medicine, University of Oxford, Oxford, United Kingdom.

³Lao-Oxford-Mahosot Hospital-Wellcome Trust Research Unit, Microbiology Laboratory, Mahosot Hospital, Vientiane, Lao People's Democratic Republic.

⁴Cambodia Oxford Medical Research Unit, Angkor Hospital for Children, Siem Reap, Cambodia. ⁵Present address: Department of Infection, University Hospital Southampton, Southampton, UK. ⁶Faculty of Infectious and Tropical Diseases, London School of Hygiene and Tropical Medicine, London, UK.

⁷Mahidol-Oxford Tropical Medicine Research Unit, Faculty of Tropical Medicine, Mahidol University, Bangkok, Thailand. ⁸Present address: London North West University Healthcare NHS Trust, Northwick Park Hospital, London, UK. ⁹Myanmar Oxford Clinical Research Unit, Yangon, Myanmar.

Received: 29 June 2020 Accepted: 4 February 2021

Published online: 12 February 2021

References

- World Health Organisation. Global antimicrobial resistance surveillance system: manual for early implementation. Geneva: World Health Organisation; 2015.
- Schneiderhan W, Grundt A, Wörner S, Findeisen P, Neumaier M. Work flow analysis of around-the-clock processing of blood culture samples and integrated MALDI-TOF mass spectrometry analysis for the diagnosis of bloodstream infections. *Clin Chem*. 2013;59:1649–56. <https://doi.org/10.1373/clinchem.2012.198218>.
- Wattal C, Oberoi JK. Microbial identification and automated antibiotic susceptibility testing directly from positive blood cultures using MALDI-TOF MS and VITEK 2. *Eur J Clin Microbiol Infect Dis*. 2016;35:75–82. <https://doi.org/10.1007/s10096-015-2510-y>.
- Satta G, Edmonstone J. Consolidation of pathology services in England: have savings been achieved? *BMC health Serv res*; 2018.
- Pentella M, Weinstein MP, Beekmann SE, Polgreen PM, Ellison RT. Impact of changes in clinical microbiology laboratory location and ownership on the practice of infectious diseases. *J Clin Microbiol*. 2020;58(5):e01508–19.
- Venturelli C, Righi E, Borsari L, Aggazzotti G, Busani S, Mussini C, et al. Impact of pre-analytical time on the recovery of pathogens from blood cultures: results from a large retrospective survey. *PLoS One*. 2017;12:e0169466. <https://doi.org/10.1371/journal.pone.0169466>.
- Sautter RL, Bills AR, Lang DL, Ruschell G, Heiter BJ, Bourbeau PP. Effects of delayed-entry conditions on the recovery and detection of microorganisms from BacT/ALERT and BACTEC blood culture bottles. *J Clin Microbiol*. 2006;44:1245–9. <https://doi.org/10.1128/JCM.44.4.1245-1249.2006>.
- Miles AA, Misra SS, Irwin JO. The estimation of the bactericidal power of the blood. *J Hyg (Lond)*. 1938;38:732–49 <https://www.ncbi.nlm.nih.gov/pubmed/20475467>.
- Ombelet S, Barbé B, Affolabi D, Ronat J-B, Lompo P, Lunguya O, et al. Best practices of blood cultures in low- and middle-income countries. *Front Med*. 2019;6:131. <https://doi.org/10.3389/fmed.2019.00131>.
- Public Health England. Investigation of blood cultures (for organisms other than Mycobacterium species). *UK Stand Microbiol Investig*. 2019;37:8.2.
- Vasallo FJ, López-Miragaya I, Rodríguez A, Torres J. Apparently false-positive blood cultures due to autolyzed Streptococcus pneumoniae. *Clin Microbiol Infect*. 2000;6:688–9.
- Adeniyi-Jones CC, Stevens DL, Rasquinha ES. False no-growth blood cultures in pneumococcal pneumonia. *J Clin Microbiol*. 1980;12:572–5.
- Fischer GW, Longfield R, Hemming VG, Valdes-Dapena A, Smith LP. Pneumococcal sepsis with false-negative blood cultures. *Am J Clin Pathol*. 1982;78:348–50.
- Fischer GW, Smith LP, Hemming VG, Longfield R, Valdes Dapena AA, Lopreiato JO. Avoidance of false-negative blood culture results by rapid detection of pneumococcal antigen. *JAMA J Am Med Assoc*. 1984;252:1742–3.
- Villar H, Longo LM, Laurino GJ, Vicente A, Gutiérrez ME, Hoffman M. Autolysis of Streptococcus pneumoniae in blood cultures. *Acta Biochim Clin Latinoam*. 2005;39:151–5.
- Barenfanger J, Graham DR, Kolluri L, Sangwan G, Lawhorn J, Drake CA, et al. Decreased mortality associated with prompt gram staining of blood cultures. *Am J Clin Pathol*. 2008;130:870–6.
- Lamy B, Dargère S, Arendrup MC, Parienti JJ, Tattevin P. How to optimize the use of blood cultures for the diagnosis of bloodstream infections? A state-of-the-art. *Front Microbiol*. 2016;7:697.
- Chapin K, Lauderdale TL. Comparison of Bactec 9240 and Difco ESP blood culture systems for detection of organisms from vials whose entry was delayed. *J Clin Microbiol*. 1996;34:543–9.
- Mayxay M, Castonguay-Vanier J, Chansamouth V, Dubot-Peres A, Paris DH, Phetsouvanh R, et al. Causes of non-malarial fever in Laos: a prospective study. *Lancet Glob Heal*. 2013;1:e46–54. [https://doi.org/10.1016/S2214-109X\(13\)70008-1](https://doi.org/10.1016/S2214-109X(13)70008-1).
- Rönnerberg C, Mildt M, Ullberg M, Özenci V. Transport time for blood culture bottles: underlying factors and its consequences. *Diagn Microbiol Infect Dis*. 2013;76:286–90.
- Palmer HR, Palavecino EL, Johnson JW, Ohl CA, Williamson JC. Clinical and microbiological implications of time-to-positivity of blood cultures in patients with Gram-negative bacilli bacteremia. *Eur J Clin Microbiol Infect Dis*. 2013;32(7):955–9.
- Sliva HS, Washington JA. Optimal time for routine early subculture of blood cultures. *J Clin Microbiol*. 1980;12:445–6.
- Ganguli LA, O'Hare W, Hyde WA. Rapid detection of bacteraemia by early subculture. *J Med Microbiol*. 1984;17:311–5.
- Todd JK, Roe MH. Rapid detection of Bacteriemia by an early subculture technic. *Am J Clin Pathol*. 1975;64:694–9.
- Szymczak EG, Barr JT, Durbin WA, Goldmann DA. Evaluation of blood culture procedures in a pediatric hospital. *J Clin Microbiol*. 1979;9:88–92.
- Duraijaj E, Lyngdoh WW, Banik A, Khyriem BA, Bhattacharyya P, Rajbongshi J, et al. Comparative evaluation of conventional (manual) blood culture system and BacT/ALERT 3D (automated) blood culture system in a Tertiary care hospital. *Sch J Appl Med Sci*. 2017;5:544–9.
- Schelonka RL, Chai MK, Yoder BA, Hensley D, Brockett RM, Ascher DP. Volume of blood required to detect common neonatal pathogens. *J Pediatr*. 1996;129(2):275–8.
- Dietzman DE, Fischer GW, Schoenkech FD. Neonatal escherichia coli septicemia-bacterial counts in blood. *J Pediatr*. 1974;85(1):128–30.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

